PRECAUTIONS AND LIMITATIONS

- Do not use this test past the expiration date.
- Make sure the package is sealed and do not open it until you are ready to use the test.
- Certain health conditions, such as an ovarian cyst or ectopic pregnancy (pregnancy outside the uterus), can cause a false or irregular result.
- Procedures should be followed precisely for accurate results.
- All results should be confirmed by your healthcare provider, especially when making decisions about future medical care.

To increase your chances for a healthy pregnancy for you and your baby, you should consider the following:

- Maintain a well-balanced diet.
- Do not smoke.

The period from conception through the third month of pregnancy is a very important time. You should see your doctor regularly throughout your pregnancy for proper care and nutritional counseling.

WARNING:

According to the Surgeon General, women should not drink alcoholic beverages during pregnancy due to the risk of birth defects.

If you have any questions about the instructions or your results, call our Help Line Monday – Friday between 7:00 a.m. – 5:00 p.m. Eastern Time. Toll-Free 1-800-367-6022, or visit us at www.FirstResponse.com

FIRST RESPONSE™ PREGNANCY

Manufactured for Church & Dwight Co., Inc., Ewing, NJ 08628
FIRST RESPONSE is a trademark of Church & Dwight Co., Inc.
www.FirstResponse.com

Need Spanish? Scan below

72031511
FRIN-90126-09

FIRST RESPONSE™ PREGNANCY

Watch How To Take The Test Here

Rapid Result Pregnancy Test

When to use the FIRST RESPONSE™ Rapid Result Pregnancy Test
The FIRST RESPONSE™ Rapid Result Pregnancy Test is designed to detect hCG (human chorionic gonadotropin). **Test on the day of your missed period or any day thereafter.**

You can use the FIRST RESPONSE™ Rapid Result Pregnancy test at any time of day. You do not have to use first morning urine. However, your first morning urine generally contains the highest level of hCG, the pregnancy hormone.

PLEASE READ ALL THE INFORMATION IN THIS INSERT BEFORE PERFORMING THE TEST.

WARNINGS AND PRECAUTIONS

For in-vitro diagnostic use. Not to be taken internally. Do not use after the expiration date. Expiration date is stamped on the bottom of the carton. **Single-use device. Do not reuse.** Keep out of the reach of children. Do not disassemble the device. Do not eat any component of this device. If accidentally swallowed, seek medical attention.

STORAGE

Store the FIRST RESPONSE™ Rapid Result Pregnancy Test in a dry place below 86ºF (30ºC). Do not freeze.

BEFORE YOU BEGIN:

Have a clock or timer. After reading the insert and when you are ready to perform the test, remove the test stick from the foil wrapper and take off the **Overcap.**

INSTRUCTIONS (How to use):

STEP 1

Test In Your Urine Stream

- Hold the test stick by the **Thumb Grip** with the Absorbent Tip pointing downward and the **Result Window** facing away from your body.
- Place the **Absorbent Tip** in your urine stream for **5 seconds only**.

OR

Test In Container

- Collect your urine in a clean, dry cup and immerse the entire **Absorbent Tip** in the urine for **20 seconds only**.



STEP 2 →

INSTRUCTIONS (How to use):

STEP 2

- While still holding the stick with the **Absorbent Tip** facing downward, replace the **Overcap** and lay the stick on a flat non-absorbent surface with the **Result Window** facing up.
- You may soon see a pink color moving across the **Result Window** to indicate the test is working.

WAIT

- **Wait 1 FULL MINUTE** for the test results to develop and a result to be displayed.
- You may see a pink color moving across the **Result Window**.
- Time your test carefully. You will be able to read your results after 1 minute.

READING THE RESULTS

PREGNANT
TWO PINK LINES in the **Result Window**.
One line may be lighter than the other.

NOT PREGNANT
ONE PINK LINE in the **Result Window**.

INVALID RESULT
Test again with another FIRST RESPONSE™ Rapid Result Test Stick

AFTER TESTING

Discard all components in a waste basket. Do not reuse the test.

72031511
FRIN-90126-09

QUESTIONS AND ANSWERS

Q: How does the test work?
A: The FIRST RESPONSE™ Rapid Result Pregnancy Test detects hCG (human chorionic gonadotropin) your body makes during pregnancy. The hormone is found in urine when pregnant.

Q: When can I test with FIRST RESPONSE™ Rapid Result Pregnancy Test?
A: You can use the FIRST RESPONSE™ Rapid Result Pregnancy Test any time of the day. You do not have to use first morning urine. Generally, however, your first morning urine contains the highest level of the pregnancy hormone. You can use the FIRST RESPONSE™ Rapid Result Pregnancy Test on the day of your missed period or any day thereafter.

Q: When should I read my results?
A: The FIRST RESPONSE™ Rapid Result Pregnancy Test results should be read as soon as possible, after waiting 1 minute. A positive result will last for at least 48 hours. A negative result should not be read after 10 minutes.

Q: How accurate is the test?
A: In studies of the FIRST RESPONSE™ Rapid Result Pregnancy Test with urine samples representative of both pregnant and non-pregnant subjects, laboratory technicians obtained the correct expected result in more than 99%* of the samples.

Q: What do I do if the result shows two pink lines (Pregnant)?
A: Remember, this test is not intended to replace your healthcare professional's diagnosis. A test result of two pink lines shows that the pregnancy hormone (hCG) was detected, which normally indicates pregnancy. Consult your doctor to confirm you are pregnant. If you test positive, but think you may not be pregnant, you should check with your doctor.

Q: What do I do if the result shows one line (Not Pregnant)?
A: You may not be pregnant, or it may be too early to tell. If you do not get your period within 7 days, you should retest with another FIRST RESPONSE™ Rapid Result Pregnancy Test. It is possible that your expected period may be later than anticipated or your urine may not have had enough pregnancy hormone for the test to give a positive result. If upon retesting you get another negative result and your period has still not started, you should call your doctor.

Q: I have waited longer than 1 minute, and still there are no lines in the Result Window. Is there anything wrong?
A: Your result should appear within 1 minute on the day of your missed period. There should be at least one pink line in the **Result Window** at 5 minutes to indicate that the test is completed and you have done it correctly. If not, the test is invalid. This could be due to insufficient urine on the **Absorbent Tip** or not holding the device with the **Absorbent Tip** pointing downward while replacing the **Overcap**. You should retest with a new FIRST RESPONSE™ Rapid Result test, carefully following all directions.

Q: Do any drugs or medications affect the test?
A: Only medications containing the pregnancy hormone, human chorionic gonadotropin (hCG), can affect the result. The test should not be affected by hormone therapies containing Clomiphene citrate, alcohol, pain killers, antibiotics, or the contraceptive pill.

Q: Will the amount of liquid I drink affect the result?
A: Excessive fluid intake including water could dilute your urine, causing the amount of hCG present not to be detected properly. You should not drink too much liquid before testing.

* 99% accurate at detecting the pregnancy hormone. However, some pregnant women may not have detectable amounts of pregnancy hormone in their urine on the first day of the missed period or may have miscalculated the first day of their period.

72031511
FRIN-90126-09